## **Department of Forensic Science Breath Alcohol Section** Trouble Call Log 1/1/1990 To 3/4/2014

**Instrument Serial Number:** 

010605

04-Mar-14

04-Mar-14 Date:

Time:

11:06 PM

Remote: True Tech:

TLN (25387)

Location:

Staunton PD

Indication: OPERATOR QUESTION / COMMENT

Problem: OPERATOR QUESTION / COMMENT. INSTRUMENT FREEZES AT OPERATOR PIN SCREEN AND WILL NOT CONTINUE

WITH TEST.

Technician Response: AGENCY CONTACTED VIA PHONE. OPERATOR CHECKED KEYBOARD CONNECTION AND REATTACHED IT.

OPERATOR PERFORMED SUCCESSFUL PRACTICE TEST.

Resolution NO FURTHER ACTION REQUIRED.